



Certificate of Calibration

for

Description: *Sound Level Calibrator*
Manufacturer: *RION*
Type No.: *NC-75*
Serial No.: *34724244*

Submitted by:

Customer: *Aurecon Hong Kong Limited*
Address: *Unit 1608, 16/F, Tower B, Manulife Financial Centre,
223-231 Wai Yip Street, Kwun Tong,
Kowloon, Hong Kong*

Upon receipt for calibration, the instrument was found to be:

- Within**
 Outside

the allowable tolerance.

The test equipments used for calibration are traceable to National Standards via:

- The Government of The Hong Kong Special Administrative Region Standard & Calibration Laboratory

Date of receipt: 22 July 2024

Date of calibration: 24 July 2024

Date of NEXT calibration: 23 July 2025

Calibrated by: 
Calibration Technician

Certified by: 
Mr. Ng Yan Wa
Laboratory Manager

Date of issue: 24 July 2024



**1. Calibration Precautions:**

- The unit-under-test (UUT) was allowed to stabilize in the laboratory for over 24 hours, and switched on to warm up for over 10 minutes before the commencement of the test.
- The results presented are the mean of 3 measurements at each calibration point.

2. Calibration Specifications:

Calibration check

3. Calibration Conditions:

Air Temperature: 23.4 °C
Air Pressure: 1005 hPa
Relative Humidity: 56.7 %

4. Calibration Equipment:

| Test Equipment | Type | Serial No. | Calibration Report Number | Traceable to |
|--------------------------|------------|------------|---------------------------|--------------|
| Multifunction Calibrator | B&K 4226 | 2288467 | AV240081 | HOKLAS |
| Sound Level Meter | RION NA-28 | 30721812 | AV230128 | HOKLAS |

5. Calibration Results

5.1 Sound Pressure Level

| Nominal value dB | Accept lower level dB | Accept upper level dB | Measured value dB |
|---------------------|--------------------------|--------------------------|----------------------|
| 94.0 | 93.6 | 94.4 | 93.9 |

Note:

The values given in this certification only related to the values measured at the time of the calibration.



Certificate of Calibration

for

Description: *Sound Level Calibrator*
Manufacturer: *RION*
Type No.: *NC-75*
Serial No.: *34724245*

Submitted by:

Customer: *Aurecon Hong Kong Limited*
Address: *Unit 1608, 16/F, Tower B, Manulife Financial Centre,
223-231 Wai Yip Street, Kwun Tong,
Kowloon, Hong Kong*

Upon receipt for calibration, the instrument was found to be:

- Within**
 Outside

the allowable tolerance.

The test equipments used for calibration are traceable to National Standards via:

- The Government of The Hong Kong Special Administrative Region Standard & Calibration Laboratory

Date of receipt: 22 July 2024

Date of calibration: 24 July 2024

Date of NEXT calibration: 23 July 2025

Calibrated by: _____

Calibration Technician

Certified by: _____

Mr. Ng Yan Wa
Laboratory Manager

Date of issue: 24 July 2024



1. Calibration Precautions:

- The unit-under-test (UUT) was allowed to stabilize in the laboratory for over 24 hours, and switched on to warm up for over 10 minutes before the commencement of the test.
- The results presented are the mean of 3 measurements at each calibration point.

2. Calibration Specifications:

Calibration check

3. Calibration Conditions:

Air Temperature: 23.4 °C
Air Pressure: 1005 hPa
Relative Humidity: 56.7 %

4. Calibration Equipment:

| Test Equipment | Type | Serial No. | Calibration Report Number | Traceable to |
|--------------------------|------------|------------|---------------------------|--------------|
| Multifunction Calibrator | B&K 4226 | 2288467 | AV240081 | HOKLAS |
| Sound Level Meter | RION NA-28 | 30721812 | AV230128 | HOKLAS |

5. Calibration Results

5.1 Sound Pressure Level

| Nominal value dB | Accept lower level dB | Accept upper level dB | Measured value dB |
|---------------------|--------------------------|--------------------------|----------------------|
| 94.0 | 93.6 | 94.4 | 94.0 |

Note:

The values given in this certification only related to the values measured at the time of the calibration.





Manufacturer Calibration Certificate

The sound level meter submitted for testing successfully completed the periodic tests of IEC 61672-3. All tests are traceable in accordance with ISO/IEC 17025.

No pattern approval is available for this sound level meter configuration.

Sound Level Meter

| | | | |
|------------------------|-----------|-----|--------------|
| Manufacturer | NTi Audio | | |
| Type | XL3 | S/N | A3A-01229-F0 |
| Firmware | V1.36 | | |
| Microphone Model | M2340 | | |
| Preamplifier | MA230 | S/N | 1794 |
| Microphone Capsule | MC230A | S/N | A28290 |
| Performance class | | | |
| Customer Inventory Nr. | | | |


Customer

Date 25 July 2024

Certificate FL-24-115

Results **PASSED**
(for detailed report see next pages)

Operator


Markus Frick

NTi Audio AG • Im alten Riet 102, 9494 Schaan • Liechtenstein
info@nti-audio.com • www.nti-audio.com

Measurement equipment

Test System

| | |
|------------------|------------------------------|
| Model | NTi Audio FX100, S/No. 11094 |
| Last Calibration | 02 July 2024 |
| Cal Certificate | NTI Cal #3393 |
| Next Calibration | 02 July 2025 |

Reference Microphone

| | |
|------------------|--|
| Model | MTG MV203 S/N #2435, Mic Capsule MK202 S/No. #7313 |
| Last Calibration | 18 November 2022 |
| Cal Certificate | DAkKS-000875 |
| Next Calibration | 17 November 2024 |

Sound Calibrator

| | |
|-----------------------|----------------------------|
| Model | Norsonic 1251 S/No. #30930 |
| Reference Level | 114 dB |
| Calibration Frequency | 1000 Hz |
| Last Calibration | 08 December 2022 |
| Cal Certificate | METAS #259-19602 |
| Next Calibration | 07 December 2024 |

Environmental conditions

| | |
|-------------|---------|
| Temperature | 25.2 °C |
| Humidity | 48 % |
| Pressure | 965 hPa |

Notes

- This calibration certificate documents the traceability to national standards, which realize the units of measurement according to the international Systems of Units (SI).
- The user is obliged to have the object recalibrated at appropriate intervals.
- This calibration certificate may not be reproduced other than in full except with the permission of the issuing laboratory. Calibration certificates without signature are not valid.
- All limits listed in this report are acceptance limits in accordance with IEC61672.
- The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with the regulations of the GUM.



1. Indication at the calibration check frequency

The indication of the sound level meter at the calibration check frequency is checked by application of the sound calibrator and adjusted, if necessary, to indicate the required sound level for the environmental conditions under which the tests are performed. All levels in [dB].

| Sensitivity before calibration | Sensitivity after calibration | Meas level | Limit - | Limit + | Uncert. | Status |
|--------------------------------|-------------------------------|------------|---------|---------|---------|--------|
| 42.9 mV/Pa | 44.4 mV/Pa | 114 | 113 | 115 | 0.2 | Passed |

2. Self-generated noise

2.1 Microphone cartridge installed

The self-generated noise is measured in the most-sensitive level range as a time-averaged sound pressure level with frequency-weighting A and an averaging time of 30 seconds. All levels in [dB].

| Weighting | Meas level | Limit + | Uncert. | Status |
|-----------|------------|---------|---------|--------|
| A | 16.0 | 19.0 | 0.1 | Passed |

2.2 Microphone cartridge replaced by the capsule replacement NTI-K65-15

The self-generated noise is measured in the most-sensitive level range as a time-averaged sound pressure level for all frequency-weightings and an averaging time of 30 seconds. All levels in [dB] referenced to $S = 42 \text{ mV/Pa}$.

| Weighting | Meas level | Limit + | Uncert. | Status |
|-----------|------------|---------|---------|--------|
| A | 10.7 | 13.0 | 0.1 | Passed |
| C | 12.8 | 16.0 | 0.1 | Passed |
| Z | 18.6 | 24.0 | 0.1 | Passed |

3. Acoustic signal tests of a frequency weighting

The frequency weighting is tested for frequency-weighting A, using an acoustic test facility. The sound level meter is set to a fast time-weighted sound level in the reference level range. All levels in [dB].

| Freq. [Hz] | Gen. level | Meas level | Dev | Limit - | Limit + | Uncert. | Status |
|------------|------------|------------|------|---------|---------|---------|--------|
| 125 | 70.0 | 69.6 | -0.4 | -1.0 | 1.0 | 0.4 | Passed |
| 250 | 77.1 | 77.0 | -0.1 | -1.0 | 1.0 | 0.4 | Passed |
| 500 | 82.7 | 82.7 | 0.0 | -1.0 | 1.0 | 0.4 | Passed |
| 1000 | 86.0 | 86.1 | 0.1 | -0.7 | 0.7 | 0.4 | Passed |
| 2000 | 87.2 | 87.4 | 0.2 | -1.0 | 1.0 | 0.4 | Passed |
| 4000 | 87.0 | 87.1 | 0.1 | -1.0 | 1.0 | 0.4 | Passed |
| 8000 | 84.8 | 84.9 | 0.1 | -2.5 | 1.5 | 0.4 | Passed |

4. Electric signal tests of frequency weightings

Frequency weightings are determined relative to the response at 1 kHz using steady sinusoidal electrical input signals. The sound level meter is set to display F-time-weighted sound level in the reference level range. All available frequency weightings provided in the sound level meter are verified. All levels in [dB].

4.1 A-Weighting

| Freq. [Hz] | Gen. level | Meas level | Dev | Limit - | Limit + | Uncert. | Status |
|------------|------------|------------|------|---------|---------|---------|--------|
| 1000 | 80.0 | 80.0 | 0.0 | -0.7 | 0.7 | 0.1 | Passed |
| 63 | 106.2 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 125 | 96.1 | 79.9 | -0.1 | -1.0 | 1.0 | 0.1 | Passed |
| 250 | 88.6 | 79.9 | -0.1 | -1.0 | 1.0 | 0.1 | Passed |
| 500 | 83.2 | 79.9 | -0.1 | -1.0 | 1.0 | 0.1 | Passed |
| 2000 | 78.8 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 4000 | 79.0 | 79.9 | -0.1 | -1.0 | 1.0 | 0.1 | Passed |
| 8000 | 81.1 | 79.6 | -0.4 | -2.5 | 1.5 | 0.1 | Passed |
| 12500 | 84.3 | 79.3 | -0.7 | -2.5 | 1.5 | 0.1 | Passed |
| 16000 | 86.6 | 78.6 | -1.4 | -2.5 | 1.5 | 0.1 | Passed |

4.2 C-Weighting

| Freq. [Hz] | Gen. level | Meas level | Dev | Limit - | Limit + | Uncert. | Status |
|------------|------------|------------|------|---------|---------|---------|--------|
| 1000 | 80.0 | 80.0 | 0.0 | -0.7 | 0.7 | 0.1 | Passed |
| 63 | 80.8 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 125 | 80.2 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 250 | 80.0 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 500 | 80.0 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 2000 | 80.2 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 4000 | 80.8 | 79.9 | -0.1 | -1.0 | 1.0 | 0.1 | Passed |
| 8000 | 83.0 | 79.6 | -0.4 | -2.5 | 1.5 | 0.1 | Passed |
| 12500 | 86.2 | 79.3 | -0.7 | -2.5 | 1.5 | 0.1 | Passed |
| 16000 | 88.5 | 78.6 | -1.4 | -2.5 | 1.5 | 0.1 | Passed |

4.3 Z-Weighting

| Freq. [Hz] | Gen. level | Meas level | Dev | Limit - | Limit + | Uncert. | Status |
|------------|------------|------------|------|---------|---------|---------|--------|
| 1000 | 80.0 | 80.0 | 0.0 | -0.7 | 0.7 | 0.1 | Passed |
| 63 | 80.0 | 79.9 | -0.1 | -1.0 | 1.0 | 0.1 | Passed |
| 125 | 80.0 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 250 | 80.0 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 500 | 80.0 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 2000 | 80.0 | 79.9 | -0.1 | -1.0 | 1.0 | 0.1 | Passed |
| 4000 | 80.0 | 79.9 | -0.1 | -1.0 | 1.0 | 0.1 | Passed |
| 8000 | 80.0 | 79.8 | -0.2 | -2.5 | 1.5 | 0.1 | Passed |
| 12500 | 80.0 | 79.8 | -0.2 | -2.5 | 1.5 | 0.1 | Passed |
| 16000 | 80.0 | 79.8 | -0.2 | -2.5 | 1.5 | 0.1 | Passed |



5. Frequency and time weightings at 1kHz

While injecting a constant steady signal at the reference frequency of 1 kHz the F-time-weighted sound level, S-time-weighted sound level and time-averaged sound level are verified with frequency weighting A. Additionally the F-time-weighted sound level for frequency weightings C and Z is measured. The first measurement serves as reference and differences in the reading with respect to this first one are determined. All levels in [dB].

| Level | Exp level | Meas level | Dev | Limit - | Limit + | Uncert. | Status |
|-------|-----------|------------|------|---------|---------|---------|--------|
| LAF | 114.0 | 114.0 | 0.0 | -0.7 | 0.7 | 0.1 | Passed |
| LAS | 114.0 | 113.8 | -0.2 | -0.7 | 0.7 | 0.1 | Passed |
| LAeq | 114.0 | 114.0 | 0.0 | -0.7 | 0.7 | 0.1 | Passed |
| LCF | 114.0 | 114.0 | 0.0 | -0.7 | 0.7 | 0.1 | Passed |
| LCeq | 114.0 | 114.0 | 0.0 | -0.7 | 0.7 | 0.1 | Passed |
| LZF | 114.0 | 114.0 | 0.0 | -0.7 | 0.7 | 0.1 | Passed |
| LZeq | 114.0 | 114.0 | 0.0 | -0.7 | 0.7 | 0.1 | Passed |



6. Level linearity on the reference level range

The level linearity on the reference level range is determined by applying steady sinusoidal electrical signals at a frequency of 8 kHz with the sound level meter set for frequency-weighting A and fast time-weighting. All levels in [dB].

| Exp abs level | Meas. level | Abs dev | Abs Limit - | Abs Limit + | Exp rel level | Rel dev | Rel Limit - | Rel Limit + | Uncert. | Status |
|---------------|-------------|---------|-------------|-------------|---------------|---------|-------------|-------------|---------|--------|
| 114.0 | 114.0 | 0.0 | -0.8 | 0.8 | 0.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 119.0 | 119.0 | 0.0 | -0.8 | 0.8 | 119.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 124.0 | 124.0 | 0.0 | -0.8 | 0.8 | 124.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 129.0 | 129.0 | 0.0 | -0.8 | 0.8 | 129.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 134.0 | 134.1 | 0.1 | -0.8 | 0.8 | 134.0 | 0.1 | -0.3 | 0.3 | 0.1 | Passed |
| 135.0 | 135.1 | 0.1 | -0.8 | 0.8 | 135.1 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 136.0 | 136.1 | 0.1 | -0.8 | 0.8 | 136.1 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 114.0 | 114.0 | 0.0 | -0.8 | 0.8 | 0.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 109.0 | 109.0 | 0.0 | -0.8 | 0.8 | 109.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 104.0 | 104.0 | 0.0 | -0.8 | 0.8 | 104.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 99.0 | 99.0 | 0.0 | -0.8 | 0.8 | 99.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 94.0 | 94.0 | 0.0 | -0.8 | 0.8 | 94.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 89.0 | 89.0 | 0.0 | -0.8 | 0.8 | 89.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 84.0 | 84.0 | 0.0 | -0.8 | 0.8 | 84.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 79.0 | 79.0 | 0.0 | -0.8 | 0.8 | 79.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 74.0 | 74.0 | 0.0 | -0.8 | 0.8 | 74.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 69.0 | 69.0 | 0.0 | -0.8 | 0.8 | 69.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 64.0 | 64.0 | 0.0 | -0.8 | 0.8 | 64.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 59.0 | 59.0 | 0.0 | -0.8 | 0.8 | 59.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 54.0 | 54.0 | 0.0 | -0.8 | 0.8 | 54.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 49.0 | 49.0 | 0.0 | -0.8 | 0.8 | 49.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 44.0 | 44.0 | 0.0 | -0.8 | 0.8 | 44.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 39.0 | 39.0 | 0.0 | -0.8 | 0.8 | 39.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 34.0 | 34.0 | 0.0 | -0.8 | 0.8 | 34.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 29.0 | 29.0 | 0.0 | -0.8 | 0.8 | 29.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 28.0 | 28.0 | 0.0 | -0.8 | 0.8 | 28.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 27.0 | 27.1 | 0.1 | -0.8 | 0.8 | 27.0 | 0.1 | -0.3 | 0.3 | 0.1 | Passed |
| 26.0 | 26.1 | 0.1 | -0.8 | 0.8 | 26.1 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 25.0 | 25.1 | 0.1 | -0.8 | 0.8 | 25.1 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |



7. Level linearity including the level range control

This test is not applicable for a single-range sound level meter.

8. Toneburst response

The response of the sound level meter to short-duration signals is tested on the reference level range with 4 kHz tonebursts that start and stop at zero crossings and are extracted from steady 4 kHz sinusoidal electrical input signals. The sound level meter is set for frequency weighting A. All levels in [dB].

The continuous signal level is 123 dB.

| Burst signal | Burst duration [ms] | Exp level | Meas level | Dev | Limit - | Limit + | Uncert. | Status |
|--------------|---------------------|-----------|------------|------|---------|---------|---------|--------|
| LAF | 200 | 122.0 | 121.9 | -0.1 | -0.5 | 0.5 | 0.2 | Passed |
| LAF | 2 | 105.0 | 104.7 | -0.3 | -1.5 | 1.0 | 0.2 | Passed |
| LAF | 0.25 | 96.0 | 95.6 | -0.4 | -3.0 | 1.0 | 0.2 | Passed |
| LAS | 200 | 115.6 | 115.5 | -0.1 | -0.5 | 0.5 | 0.2 | Passed |
| LAS | 2 | 96.0 | 95.9 | -0.1 | -3.0 | 1.0 | 0.2 | Passed |
| LAeq10s | 200 | 106.0 | 105.9 | -0.1 | -0.5 | 0.5 | 0.2 | Passed |
| LAeq10s | 2 | 86.0 | 85.9 | -0.1 | -0.5 | 0.5 | 0.2 | Passed |
| LAeq10s | 0.25 | 77.0 | 76.6 | -0.4 | -0.5 | 0.5 | 0.2 | Passed |

9. C-weighted peak sound level

The sound level meter is tested on the least-sensitive level range with fast time weighting and C frequency weighting. The test signals are a single complete cycle of an 8 kHz sinusoid starting and stopping at zero crossings and positive and negative half cycles of a 500 Hz sinusoid that also start and stop at zero crossings. All levels in [dB].

| Burst signal | Source level | Exp LCp-LCF | Meas LCp-LCF | Dev | Limit - | Limit + | Uncert. | Status |
|--------------|--------------|-------------|--------------|------|---------|---------|---------|--------|
| 8kHz | 114.0 | 3.4 | 3.3 | -0.1 | -2.0 | 2.0 | 0.2 | Passed |
| 500Hz + | 132.0 | 2.4 | 2.2 | -0.2 | -1.0 | 1.0 | 0.2 | Passed |
| 500Hz - | 132.0 | 2.4 | 2.2 | -0.2 | -1.0 | 1.0 | 0.2 | Passed |

10. Overload Indication

Overload indication is tested on the least-sensitive level range with the sound level meter set to A-weighted, time-averaged sound level. Positive and negative one-half-cycle sinusoidal electrical signals at a frequency of 4 kHz are used. All levels in [dB].

| Start level | OV + | OV - | Dev | Limit - | Limit + | Uncert. | Status |
|-------------|-------|-------|-----|---------|---------|---------|--------|
| 136.5 | 139.1 | 139.1 | 0.0 | -1.5 | 1.5 | 0.3 | Passed |



Manufacturer Calibration Certificate

The sound level meter submitted for testing successfully completed the periodic tests of IEC 61672-3. All tests are traceable in accordance with ISO/IEC 17025.

No pattern approval is available for this sound level meter configuration.

Sound Level Meter

| | | | |
|------------------------|-----------|-----|--------------|
| Manufacturer | NTi Audio | | |
| Type | XL3 | S/N | A3A-01230-F0 |
| Firmware | V1.36 | | |
| Microphone Model | M2340 | | |
| Pre-amplifier | MA230 | S/N | 1797 |
| Microphone Capsule | MC230A | S/N | A28287 |
| Performance class | | | |
| Customer Inventory Nr. | | | |


Customer

Date 25 July 2024

Certificate FL-24-114

Results **PASSED**
(for detailed report see next pages)

Operator


Markus Frick

NTi Audio AG • Im alten Riet 102, 9494 Schaan • Liechtenstein
info@nti-audio.com • www.nti-audio.com

Measurement equipment

Test System

| | |
|------------------|------------------------------|
| Model | NTi Audio FX100, S/No. 11094 |
| Last Calibration | 02 July 2024 |
| Cal Certificate | NTI Cal #3393 |
| Next Calibration | 02 July 2025 |

Reference Microphone

| | |
|------------------|--|
| Model | MTG MV203 S/N #2435, Mic Capsule MK202 S/No. #7313 |
| Last Calibration | 18 November 2022 |
| Cal Certificate | DAkKS-000875 |
| Next Calibration | 17 November 2024 |

Sound Calibrator

| | |
|-----------------------|----------------------------|
| Model | Norsonic 1251 S/No. #30930 |
| Reference Level | 114 dB |
| Calibration Frequency | 1000 Hz |
| Last Calibration | 08 December 2022 |
| Cal Certificate | METAS #259-19602 |
| Next Calibration | 07 December 2024 |

Environmental conditions

| | |
|-------------|---------|
| Temperature | 23.6 °C |
| Humidity | 53 % |
| Pressure | 965 hPa |

Notes

- This calibration certificate documents the traceability to national standards, which realize the units of measurement according to the international Systems of Units (SI).
- The user is obliged to have the object recalibrated at appropriate intervals.
- This calibration certificate may not be reproduced other than in full except with the permission of the issuing laboratory. Calibration certificates without signature are not valid.
- All limits listed in this report are acceptance limits in accordance with IEC61672.
- The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with the regulations of the GUM.

1. Indication at the calibration check frequency

The indication of the sound level meter at the calibration check frequency is checked by application of the sound calibrator and adjusted, if necessary, to indicate the required sound level for the environmental conditions under which the tests are performed. All levels in [dB].

| Sensitivity before calibration | Sensitivity after calibration | Meas level | Limit - | Limit + | Uncert. | Status |
|--------------------------------|-------------------------------|------------|---------|---------|---------|--------|
| 42.8 mV/Pa | 42.9 mV/Pa | 114 | 113 | 115 | 0.2 | Passed |

2. Self-generated noise

2.1 Microphone cartridge installed

The self-generated noise is measured in the most-sensitive level range as a time-averaged sound pressure level with frequency-weighting A and an averaging time of 30 seconds. All levels in [dB].

| Weighting | Meas level | Limit + | Uncert. | Status |
|-----------|------------|---------|---------|--------|
| A | 16.2 | 19.0 | 0.1 | Passed |

2.2 Microphone cartridge replaced by the capsule replacement NTI-K65-15

The self-generated noise is measured in the most-sensitive level range as a time-averaged sound pressure level for all frequency-weightings and an averaging time of 30 seconds. All levels in [dB] referenced to S = 42 mV/Pa.

| Weighting | Meas level | Limit + | Uncert. | Status |
|-----------|------------|---------|---------|--------|
| A | 10.7 | 13.0 | 0.1 | Passed |
| C | 12.7 | 16.0 | 0.1 | Passed |
| Z | 18.7 | 24.0 | 0.1 | Passed |

3. Acoustic signal tests of a frequency weighting

The frequency weighting is tested for frequency-weighting A, using an acoustic test facility. The sound level meter is set to a fast time-weighted sound level in the reference level range. All levels in [dB].

| Freq. [Hz] | Gen. level | Meas level | Dev | Limit - | Limit + | Uncert. | Status |
|------------|------------|------------|------|---------|---------|---------|--------|
| 125 | 70.0 | 69.6 | -0.4 | -1.0 | 1.0 | 0.4 | Passed |
| 250 | 77.1 | 76.9 | -0.2 | -1.0 | 1.0 | 0.4 | Passed |
| 500 | 82.7 | 82.9 | 0.2 | -1.0 | 1.0 | 0.4 | Passed |
| 1000 | 86.0 | 86.2 | 0.2 | -0.7 | 0.7 | 0.4 | Passed |
| 2000 | 87.2 | 87.5 | 0.3 | -1.0 | 1.0 | 0.4 | Passed |
| 4000 | 87.0 | 87.2 | 0.2 | -1.0 | 1.0 | 0.4 | Passed |
| 8000 | 84.8 | 85.0 | 0.2 | -2.5 | 1.5 | 0.4 | Passed |

4. Electric signal tests of frequency weightings

Frequency weightings are determined relative to the response at 1 kHz using steady sinusoidal electrical input signals. The sound level meter is set to display F-time-weighted sound level in the reference level range. All available frequency weightings provided in the sound level meter are verified. All levels in [dB].

4.1 A-Weighting

| Freq. [Hz] | Gen. level | Meas level | Dev | Limit - | Limit + | Uncert. | Status |
|------------|------------|------------|------|---------|---------|---------|--------|
| 1000 | 80.0 | 80.0 | 0.0 | -0.7 | 0.7 | 0.1 | Passed |
| 63 | 106.2 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 125 | 96.1 | 79.9 | -0.1 | -1.0 | 1.0 | 0.1 | Passed |
| 250 | 88.6 | 79.9 | -0.1 | -1.0 | 1.0 | 0.1 | Passed |
| 500 | 83.2 | 79.9 | -0.1 | -1.0 | 1.0 | 0.1 | Passed |
| 2000 | 78.8 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 4000 | 79.0 | 79.9 | -0.1 | -1.0 | 1.0 | 0.1 | Passed |
| 8000 | 81.1 | 79.6 | -0.4 | -2.5 | 1.5 | 0.1 | Passed |
| 12500 | 84.3 | 79.3 | -0.7 | -2.5 | 1.5 | 0.1 | Passed |
| 16000 | 86.6 | 78.6 | -1.4 | -2.5 | 1.5 | 0.1 | Passed |

4.2 C-Weighting

| Freq. [Hz] | Gen. level | Meas level | Dev | Limit - | Limit + | Uncert. | Status |
|------------|------------|------------|------|---------|---------|---------|--------|
| 1000 | 80.0 | 80.0 | 0.0 | -0.7 | 0.7 | 0.1 | Passed |
| 63 | 80.8 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 125 | 80.2 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 250 | 80.0 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 500 | 80.0 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 2000 | 80.2 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 4000 | 80.8 | 79.9 | -0.1 | -1.0 | 1.0 | 0.1 | Passed |
| 8000 | 83.0 | 79.6 | -0.4 | -2.5 | 1.5 | 0.1 | Passed |
| 12500 | 86.2 | 79.3 | -0.7 | -2.5 | 1.5 | 0.1 | Passed |
| 16000 | 88.5 | 78.6 | -1.4 | -2.5 | 1.5 | 0.1 | Passed |

4.3 Z-Weighting

| Freq. [Hz] | Gen. level | Meas level | Dev | Limit - | Limit + | Uncert. | Status |
|------------|------------|------------|------|---------|---------|---------|--------|
| 1000 | 80.0 | 80.0 | 0.0 | -0.7 | 0.7 | 0.1 | Passed |
| 63 | 80.0 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 125 | 80.0 | 80.1 | 0.1 | -1.0 | 1.0 | 0.1 | Passed |
| 250 | 80.0 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 500 | 80.0 | 80.1 | 0.1 | -1.0 | 1.0 | 0.1 | Passed |
| 2000 | 80.0 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 4000 | 80.0 | 80.0 | 0.0 | -1.0 | 1.0 | 0.1 | Passed |
| 8000 | 80.0 | 79.9 | -0.1 | -2.5 | 1.5 | 0.1 | Passed |
| 12500 | 80.0 | 79.9 | -0.1 | -2.5 | 1.5 | 0.1 | Passed |
| 16000 | 80.0 | 79.9 | -0.1 | -2.5 | 1.5 | 0.1 | Passed |



5. Frequency and time weightings at 1kHz

While injecting a constant steady signal at the reference frequency of 1 kHz the F-time-weighted sound level, S-time-weighted sound level and time-averaged sound level are verified with frequency weighting A. Additionally the F-time-weighted sound level for frequency weightings C and Z is measured. The first measurement serves as reference and differences in the reading with respect to this first one are determined. All levels in [dB].

| Level | Exp level | Meas level | Dev | Limit - | Limit + | Uncert. | Status |
|-------|-----------|------------|------|---------|---------|---------|--------|
| LAF | 114.0 | 114.0 | 0.0 | -0.7 | 0.7 | 0.1 | Passed |
| LAS | 114.0 | 113.8 | -0.2 | -0.7 | 0.7 | 0.1 | Passed |
| LAeq | 114.0 | 114.0 | 0.0 | -0.7 | 0.7 | 0.1 | Passed |
| LCF | 114.0 | 114.0 | 0.0 | -0.7 | 0.7 | 0.1 | Passed |
| LCeq | 114.0 | 114.0 | 0.0 | -0.7 | 0.7 | 0.1 | Passed |
| LZF | 114.0 | 114.0 | 0.0 | -0.7 | 0.7 | 0.1 | Passed |
| LZeq | 114.0 | 114.0 | 0.0 | -0.7 | 0.7 | 0.1 | Passed |

6. Level linearity on the reference level range

The level linearity on the reference level range is determined by applying steady sinusoidal electrical signals at a frequency of 8 kHz with the sound level meter set for frequency-weighting A and fast time-weighting. All levels in [dB].

| Exp abs level | Meas. level | Abs dev | Abs Limit - | Abs Limit + | Exp rel level | Rel dev | Rel Limit - | Rel Limit + | Uncert. | Status |
|---------------|-------------|---------|-------------|-------------|---------------|---------|-------------|-------------|---------|--------|
| 114.0 | 114.0 | 0.0 | -0.8 | 0.8 | 0.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 119.0 | 119.0 | 0.0 | -0.8 | 0.8 | 119.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 124.0 | 124.0 | 0.0 | -0.8 | 0.8 | 124.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 129.0 | 129.0 | 0.0 | -0.8 | 0.8 | 129.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 134.0 | 134.0 | 0.0 | -0.8 | 0.8 | 134.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 135.0 | 135.0 | 0.0 | -0.8 | 0.8 | 135.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 136.0 | 136.0 | 0.0 | -0.8 | 0.8 | 136.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 114.0 | 114.0 | 0.0 | -0.8 | 0.8 | 0.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 109.0 | 109.0 | 0.0 | -0.8 | 0.8 | 109.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 104.0 | 104.0 | 0.0 | -0.8 | 0.8 | 104.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 99.0 | 99.0 | 0.0 | -0.8 | 0.8 | 99.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 94.0 | 94.0 | 0.0 | -0.8 | 0.8 | 94.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 89.0 | 89.0 | 0.0 | -0.8 | 0.8 | 89.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 84.0 | 84.0 | 0.0 | -0.8 | 0.8 | 84.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 79.0 | 79.0 | 0.0 | -0.8 | 0.8 | 79.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 74.0 | 74.0 | 0.0 | -0.8 | 0.8 | 74.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 69.0 | 69.0 | 0.0 | -0.8 | 0.8 | 69.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 64.0 | 64.0 | 0.0 | -0.8 | 0.8 | 64.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 59.0 | 59.0 | 0.0 | -0.8 | 0.8 | 59.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 54.0 | 54.0 | 0.0 | -0.8 | 0.8 | 54.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 49.0 | 49.0 | 0.0 | -0.8 | 0.8 | 49.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 44.0 | 44.0 | 0.0 | -0.8 | 0.8 | 44.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 39.0 | 39.0 | 0.0 | -0.8 | 0.8 | 39.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 34.0 | 34.0 | 0.0 | -0.8 | 0.8 | 34.0 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 29.0 | 29.1 | 0.1 | -0.8 | 0.8 | 29.0 | 0.1 | -0.3 | 0.3 | 0.1 | Passed |
| 28.0 | 28.0 | 0.0 | -0.8 | 0.8 | 28.1 | -0.1 | -0.3 | 0.3 | 0.1 | Passed |
| 27.0 | 27.1 | 0.1 | -0.8 | 0.8 | 27.0 | 0.1 | -0.3 | 0.3 | 0.1 | Passed |
| 26.0 | 26.1 | 0.1 | -0.8 | 0.8 | 26.1 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |
| 25.0 | 25.1 | 0.1 | -0.8 | 0.8 | 25.1 | 0.0 | -0.3 | 0.3 | 0.1 | Passed |

7. Level linearity including the level range control

This test is not applicable for a single-range sound level meter.

8. Toneburst response

The response of the sound level meter to short-duration signals is tested on the reference level range with 4 kHz tonebursts that start and stop at zero crossings and are extracted from steady 4 kHz sinusoidal electrical input signals. The sound level meter is set for frequency weighting A. All levels in [dB].

The continuous signal level is 123 dB.

| Burst signal | Burst duration [ms] | Exp level | Meas level | Dev | Limit - | Limit + | Uncert. | Status |
|--------------|---------------------|-----------|------------|------|---------|---------|---------|--------|
| LAF | 200 | 122.0 | 121.9 | -0.1 | -0.5 | 0.5 | 0.2 | Passed |
| LAF | 2 | 105.0 | 104.8 | -0.2 | -1.5 | 1.0 | 0.2 | Passed |
| LAF | 0.25 | 96.0 | 95.6 | -0.4 | -3.0 | 1.0 | 0.2 | Passed |
| LAS | 200 | 115.6 | 115.5 | -0.1 | -0.5 | 0.5 | 0.2 | Passed |
| LAS | 2 | 96.0 | 95.9 | -0.1 | -3.0 | 1.0 | 0.2 | Passed |
| LAeq10s | 200 | 106.0 | 105.8 | -0.2 | -0.5 | 0.5 | 0.2 | Passed |
| LAeq10s | 2 | 86.0 | 85.9 | -0.1 | -0.5 | 0.5 | 0.2 | Passed |
| LAeq10s | 0.25 | 77.0 | 76.8 | -0.2 | -0.5 | 0.5 | 0.2 | Passed |

9. C-weighted peak sound level

The sound level meter is tested on the least-sensitive level range with fast time weighting and C frequency weighting. The test signals are a single complete cycle of an 8 kHz sinusoid starting and stopping at zero crossings and positive and negative half cycles of a 500 Hz sinusoid that also start and stop at zero crossings. All levels in [dB].

| Burst signal | Source level | Exp LCp-LCF | Meas LCp-LCF | Dev | Limit - | Limit + | Uncert. | Status |
|--------------|--------------|-------------|--------------|------|---------|---------|---------|--------|
| 8kHz | 114.0 | 3.4 | 3.2 | -0.2 | -2.0 | 2.0 | 0.2 | Passed |
| 500Hz + | 132.0 | 2.4 | 2.2 | -0.2 | -1.0 | 1.0 | 0.2 | Passed |
| 500Hz - | 132.0 | 2.4 | 2.2 | -0.2 | -1.0 | 1.0 | 0.2 | Passed |

10. Overload Indication

Overload indication is tested on the least-sensitive level range with the sound level meter set to A-weighted, time-averaged sound level. Positive and negative one-half-cycle sinusoidal electrical signals at a frequency of 4 kHz are used. All levels in [dB].

| Start level | OV + | OV - | Dev | Limit - | Limit + | Uncert. | Status |
|-------------|-------|-------|-----|---------|---------|---------|--------|
| 136.8 | 139.4 | 139.5 | 0.1 | -1.5 | 1.5 | 0.3 | Passed |