



Acumen Laboratory and Testing Limited

Workshop 04, 7/F, The Whitney, No. 183 Wai Yip Street, Kwun Tong, Kowloon
Tel: (852) 2333 6823 Fax: (852) 2333 1316

Appendix - Quality Control Summary Table

Project Name: Service Contract No. WD/03/2023 Hung Shui Kiu/Ha Tsuen New Development Area Second Phase Development - Environmental Team

Sampling Date	Job No.	Unit	Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
			MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
			mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
02/12/2024	R242589		0.22	0.10	4.77	4.94	-3.50	10	93.3	Pass
04/12/2024	R242611		0.22	0.07	3.38	3.24	4.23	10	94.3	Pass
06/12/2024	R242631		0.22	0.08	3.64	3.78	-3.77	10	94.0	Pass
09/12/2024	R242636		0.22	0.09	3.31	3.25	1.83	10	94.5	Pass
11/12/2024	R242651		0.22	0.09	4.16	4.01	3.67	10	93.7	Pass
13/12/2024	R242667		0.22	0.11	3.58	3.64	-1.66	10	92.8	Pass
16/12/2024	R242690		0.22	0.09	4.57	4.38	4.35	10	93.8	Pass
18/12/2024	R242694		0.22	0.07	4.15	4.30	-3.55	10	94.0	Pass
20/12/2024	R242712		0.22	0.07	3.50	3.63	-3.65	10	94.4	Pass
23/12/2024	R242738		0.22	0.11	4.18	3.99	4.65	10	93.0	Pass
25/12/2024	R242739		0.22	0.09	3.65	3.49	4.48	10	93.8	Pass
27/12/2024	R242754		0.22	0.09	3.37	3.48	-3.21	10	94.8	Pass
30/12/2024	R242788		0.22	0.07	4.15	4.01	3.43	10	93.1	Pass



Acumen Laboratory and Testing Limited

Flat/Rm D, 12/F, Ford Glory Plaza, Nos. 37-39 Wing Hong Street, Cheung Sha Wan, Kowloon, Hong Kong
 Tel: (852) 2333 6823 Fax: (852) 2333 1316

Appendix - Quality Control Summary Table

Project Name: Service Contract No. WD/03/2023 Hung Shui Kiu/Ha Tsuen New Development Area Second Phase Development - Environmental Team

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. / Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
02/12/2024	R242595	0.22	0.08	3.55	3.69	-3.87	10	93.3	Pass
04/12/2024	R242615	0.22	0.09	3.40	3.30	2.99	10	93.6	Pass
06/12/2024	R242633	0.22	0.08	3.35	3.46	-3.23	10	92.5	Pass
09/12/2024	R242638	0.22	0.07	3.13	3.03	3.25	10	94.2	Pass
11/12/2024	R242649	0.22	0.08	4.69	4.84	-3.15	10	93.8	Pass
13/12/2024	R242666	0.22	0.11	4.33	4.16	4.00	10	92.3	Pass
16/12/2024	R242689	0.22	0.08	3.02	3.11	-2.94	10	93.5	Pass
18/12/2024	R242696	0.22	0.10	3.75	3.60	4.08	10	93.3	Pass
20/12/2024	R242726	0.22	0.09	3.51	3.65	-3.94	10	93.5	Pass
23/12/2024	R242737	0.22	0.11	4.04	4.18	-3.41	10	93.8	Pass
27/12/2024	R242755	0.22	0.07	4.64	4.47	3.73	10	94.4	Pass
30/12/2024	R242790	0.22	0.11	3.95	4.08	-3.24	10	94.1	Pass