



Acumen Laboratory and Testing Limited

Flat/Rm D, 12/F, Ford Glory Plaza, Nos. 37-39 Wing Hong Street, Cheung Sha Wan, Kowloon, Hong Kong
Tel: (852) 2333 6823 Fax: (852) 2333 1316

Page 1 of 1

Appendix - Quality Control Summary Table

Project Name: Service Contract No. WD/03/2023 Hung Shui Kiu/Ha Tsuen New Development Area Second Phase Development -
Environmental Team

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
01/02/2025	R250195	0.22	0.09	5.21	5.01	3.91	10	93.0	Pass
03/02/2025	R250201	0.22	0.07	4.54	4.69	-3.25	10	94.6	Pass
05/02/2025	R250235	0.22	0.08	5.04	4.86	3.64	10	92.5	Pass
08/02/2025	R250258	0.22	0.09	3.39	3.45	-1.75	10	93.4	Pass
10/02/2025	R250264	0.22	0.09	4.06	3.96	2.49	10	93.7	Pass
12/02/2025	R250305	0.22	0.07	3.83	3.92	-2.32	10	93.3	Pass
14/02/2025	R250332	0.22	0.10	3.86	3.73	3.43	10	93.8	Pass
17/02/2025	R250335	0.22	0.08	3.33	3.44	-3.25	10	94.4	Pass
20/02/2025	R250367	0.22	0.09	3.08	2.98	3.30	10	93.8	Pass
22/02/2025	R250391	0.22	0.07	3.11	3.23	-3.79	10	93.8	Pass
24/02/2025	R250397	0.22	0.10	3.62	3.51	3.09	10	93.9	Pass
26/02/2025	R250412	0.22	0.10	3.77	3.87	-2.62	10	95.5	Pass
28/02/2025	R250436	0.22	0.09	3.37	3.26	3.32	10	94.5	Pass



Acumen Laboratory and Testing Limited

Workshop 04, 7/F, The Whitney, No. 183 Wai Yip Street, Kwun Tong, Kowloon

Tel: (852) 2333 6823 Fax: (852) 2333 1316

Page 1 of 1

Appendix - Quality Control Summary Table

Project Name: Service Contract No. WD/03/2023 Hung Shui Kiu/Ha Tsuen New Development Area Second Phase Development - Environmental Team)

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
01/02/2025	R250194	0.22	0.09	4.17	4.01	3.91	10	93.2	Pass
03/02/2025	R250200	0.22	0.11	4.83	5.01	-3.66	10	94.0	Pass
05/02/2025	R250234	0.22	0.09	4.78	4.65	2.76	10	92.8	Pass
08/02/2025	R250257	0.22	0.10	4.56	4.65	-1.95	10	92.5	Pass
10/02/2025	R250263	0.22	0.10	4.69	4.52	3.69	10	94.6	Pass
12/02/2025	R250304	0.22	0.07	4.45	4.26	4.36	10	94.5	Pass
14/02/2025	R250331	0.22	0.07	3.73	3.85	-3.17	10	92.2	Pass
17/02/2025	R250334	0.22	0.08	5.04	4.93	2.21	10	93.5	Pass
20/02/2025	R250366	0.22	0.09	4.56	4.71	-3.24	10	95.0	Pass
22/02/2025	R250390	0.22	0.10	3.55	3.42	3.73	10	94.1	Pass
24/02/2025	R250396	0.22	0.08	5.17	5.37	-3.80	10	93.9	Pass
26/02/2025	R250411	0.22	0.10	4.32	4.21	2.58	10	95.1	Pass
28/02/2025	R250435	0.22	0.09	3.60	3.74	-3.81	10	93.0	Pass