



Acumen Laboratory and Testing Limited

Workshop 04, 7/F, The Whitney, No. 183 Wai Yip Street, Kwun Tong, Kowloon

Tel: (852) 2333 6823 Fax: (852) 2333 1316

Page 1 of 1

Appendix - Quality Control Summary Table

Project Name: Service Contract No. WD/03/2023 Hung Shui Kiu/Ha Tsuen New Development Area Second Phase Development -
Environmental Team

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
02/06/2025	R251276	0.22	0.08	5.05	5.14	-1.77	10	93.2	Pass
04/06/2025	R251298	0.22	0.09	4.97	4.78	3.90	10	92.9	Pass
06/06/2025	R251313	0.22	0.08	4.57	4.69	-2.59	10	93.8	Pass
09/06/2025	R251333	0.22	0.08	3.38	3.29	2.70	10	93.2	Pass
11/06/2025	R251339	0.22	0.09	3.44	3.49	-1.44	10	93.3	Pass
14/06/2025	R251347	0.22	0.08	3.62	3.50	3.37	10	93.2	Pass
16/06/2025	R251362	0.22	0.09	4.26	4.08	4.32	10	94.2	Pass
18/06/2025	R251390	0.22	0.11	3.75	3.81	-1.59	10	94.3	Pass
21/06/2025	R251418	0.22	0.07	3.49	3.37	3.50	10	93.8	Pass
23/06/2025	R251443	0.22	0.11	3.49	3.61	-3.38	10	93.3	Pass
26/06/2025	R251472	0.22	0.09	4.10	4.00	2.47	10	93.1	Pass
28/06/2025	R251508	0.22	0.10	3.10	3.17	-2.23	10	92.8	Pass
30/06/2025	R251515	0.22	0.07	5.03	4.83	4.06	10	93.2	Pass



Acumen Laboratory and Testing Limited

Flat/Rm D, 12/F, Ford Glory Plaza, Nos. 37-39 Wing Hong Street, Cheung Sha Wan, Kowloon, Hong Kong

Tel: (852) 2333 6823 Fax: (852) 2333 1316

Page 1 of 1

Appendix - Quality Control Summary Table

Project Name: Service Contract No. WD/03/2023 Hung Shui Kiu/Ha Tsuen New Development Area Second Phase Development -
Environmental Team

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
02/06/2025	R251277	0.22	0.10	4.91	4.79	2.47	10	93.7	Pass
04/06/2025	R251299	0.22	0.07	3.46	3.51	-1.43	10	94.5	Pass
06/06/2025	R251314	0.22	0.07	4.11	3.99	2.96	10	95.5	Pass
09/06/2025	R251334	0.22	0.07	4.60	4.74	-3.00	10	94.0	Pass
11/06/2025	R251340	0.22	0.08	4.63	4.49	3.07	10	93.5	Pass
14/06/2025	R251348	0.22	0.08	4.00	4.10	-2.47	10	93.3	Pass
16/06/2025	R251363	0.22	0.11	5.05	4.91	2.81	10	94.3	Pass
18/06/2025	R251391	0.22	0.09	4.76	4.90	-2.90	10	94.6	Pass
21/06/2025	R251419	0.22	0.08	3.63	3.50	3.65	10	93.6	Pass
23/06/2025	R251444	0.22	0.10	3.08	3.18	-3.19	10	92.9	Pass
26/06/2025	R251473	0.22	0.07	5.11	4.91	3.99	10	93.9	Pass
28/06/2025	R251509	0.22	0.07	3.94	4.11	-4.22	10	93.5	Pass
30/06/2025	R251516	0.22	0.09	4.36	4.20	3.74	10	93.1	Pass